

Stretch Vinyl Exam Gloves



Durable & comfortable, these stretch vinyl gloves are manufactured with a polyvinyl chloride formulation - serving as a great alternative for latex-sensitive individuals.

- Latex Free
- Powder Free
- Beaded Cuff
- DEHP/DOP Free

Glove Thickness

Cuff - 3.2mil

Palm - 4.4mil

Fingertip - 3.6mil



Ordering Information



Item #	Concordance #	Description	UOM
HCS5500	191299	Glove, Exam, Stretch Vinyl, Small	100/Bx
HCS5501	191300	Glove, Exam, Stretch Vinyl, Medium	100/Bx
HCS5502	191301	Glove, Exam, Stretch Vinyl, Large	100/Bx
HCS5503	191302	Glove, Exam, Stretch Vinyl, XLarge	100/Bx

hcsmed.com

Contact your local Concordance Healthcare Solutions Account Manager

Rev. 014.042219

Distributed by:





Technical Information: Synthetic Stretch Vinyl Exam Gloves

Standards of Conformance:

Standard	Description
ASTM D -5250-92	Standard Specification for Polyvinyl Chloride Examination Gloves
ASTM D-5151	Standard Test Method for the Detection of Holes in Medical Gloves
ASTM D-61874	Standard Test Method for Residual Powder on Medical Gloves
ASTM D-5712	Standard Test Method for Analysis of Protein in Natural Rubber and its Products
EN 455-1	Medical Gloves for Single Use Requirements & Testing for Freedom of Holes
EN 455-2	Medical Gloves for Single Use Requirements & Testing for Physical Properties.
MIL-STD-105E (ISO 2859)	Sampling Procedures and Tables

Physical Properties (ASTM Requirements)

Parameter	Unaged	Aged
	ASTM D6319-00a Specification	ASTM D6319-00a Specification
Tensile Strength, MPa	11 MPa	11 MPa
Elongation %	300%	300%
Freedom from Holes	2.5 AQL*	N/A

*Acceptable quality level - maximum % of gloves with defects